

Issued by Notified Body No. 0402 according to Regulation (EU) 2016/425 annex V, regarding:

## Filtering half masks to protect against particles

Issued to

**Airinum AB**

Östermalmsgatan 26A, 114 26 Stockholm, Sweden

### Product description and product name

Filtering half masks to protect against particles, complying with EN 149:2001+A1:2009.

Product name: *Air Mask Lite*. For details, see appendix.

### Technical documentation

The manufacturer's technical file, latest dated 2022-04-27.

### Certificate

RISE Research Institutes of Sweden AB, Notified Body no. 0402, hereby certifies that the manufacturer's technical file and the product have been inspected in accordance with the procedure described in annex V to regulation (EU) 2016/425 and found to fulfil the requirements. The product is judged to belong to category III according to annex I to the regulation. It is a prerequisite that manufactured products, or the manufacturing process must be inspected in accordance with annex VII or VIII of the Regulation (EU) 2016/425. The EU Type-Examination is based on a type test in accordance with EN 149:2001+A1:2009. RISE Certification Rule SPCR 309 issue 2021-12-21 has been applied.

### Miscellaneous

The manufacturer's information in English on use, cleaning, storage etc. has been inspected and found to fulfil the relevant requirements of the regulation. This information must be available in the language used where the product is to be sold.

### Marking

The product shall be marked: "FFP2 R D".

### Validity

The certificate remains valid as long as the conditions laid down in the harmonized standard(s) in reference are not modified significantly or at the latest until 2027-04-29. The validity of this certificate can be verified by RISE.



Martin Tillander

Certificate C900554 | issue 2 | 2022-05-01

**RISE Research Institutes of Sweden AB | Certification**

Box 857, SE-50115 Borås, Sweden

+46 10 516 50 00 | [certifiering@ri.se](mailto:certifiering@ri.se) | [www.ri.se](http://www.ri.se)

178592

This document is the property of RISE and may not be reproduced other than in full, except with the prior written approval by RISE


### Product information

Air Mask Lite including replaceable filter Air Filter Lite – Opti, comes in 10 different colours and 3 different sizes as specified below.

Photo	Colour	Model No / Size
	Storm Black	AML-1201 / S AML-1301 / M AML-1401 / L
	Misty Grey	AML-1202 / S AML-1302 / M AML-1402 / L
	Dusty Pink	AML-1203 / S AML-1303 / M AML-1403 / L
	Sand Beige	AML-1204 / S AML-1304 / M AML-1404 / L
	Glacier White	AML-1205 / S AML-1305 / M AML-1405 / L
	Earth Tone	AML-1206 / S AML-1306 / M AML-1406 / L

Photo	Colour	Model No / Size
	Ocean Blue	AML-1207 / S AML-1307 / M AML-1407 / L
	Glacier Ice	AML-1208 / S AML-1308 / M AML-1408 / L
	Petrol Blue	AML-1210 / S AML-1310 / M AML-1410 / L
	Smog Grey	AML-1211 / S AML-1311 / M AML-1411 / L


Replaceable filter Air Filter Lite – Opti comes in 3 different sizes

Photo of Air Filter Lite - Opti	Model No / Size
	AFL-1202-3 / S AFL-1302-3 / M AFL-1402-3 / L

### Marking



Picture of marking on mask skin Air Mask Lite



Picture of marking on Air Filter Lite - Opti

### Performance:

Filtering class: FFP2

Reusable: R

Clogging performance: Yes

### Certificate history

Date	Issue	
2022-04-29	1	Certificate issued
2022-05-01	2	Certificate corrected.

Certificate C900554 | issue 2 | 2022-05-01

RISE Research Institutes of Sweden AB | Certification